



TEST REPORT

Patient ID 01040532280 Sample ID V. Testor Test Time 2023-02-21 13:32

Detected	+ SARS-CoV-2
-----------------	----------------------------

User	administrator	Test Status	✓ Completed
		Internal Controls	✓ Passed

RESULT DETAILS		Ct / EP


Viruses	Not detected	Adenovirus	- / -
	Not detected	Bocavirus	- / -
	Not detected	Coronavirus 229E	- / -
	Not detected	Coronavirus HKU1	- / -
	Not detected	Coronavirus NL63	- / -
	Not detected	Coronavirus OC43	- / -
	Not detected	Human Metapneumovirus A+B	- / -
	Not detected	Influenza A	- / -
	Not detected	Influenza A H1	- / -
	Not detected	Influenza A H1N1 pdm09	- / -
	Not detected	Influenza A H3	- / -
	Not detected	Influenza B	- / -
	Not detected	Parainfluenza virus 1	- / -
	Not detected	Parainfluenza virus 2	- / -
	Not detected	Parainfluenza virus 3	- / -
	Not detected	Parainfluenza virus 4	- / -
	Not detected	Respiratory Syncytial Virus A+B	- / -
	Not detected	Rhinovirus/Enterovirus	- / -
+	Detected	SARS-CoV-2	19.6 / 353,927
Bacteria	Not detected	<i>Bordetella pertussis</i>	- / -
	Not detected	<i>Chlamydophila pneumoniae</i>	- / -
	Not detected	<i>Legionella pneumophila</i>	- / -
	Not detected	<i>Mycoplasma pneumoniae</i>	- / -
Controls	+ Detected	IC	32.3 / 176,665

TEST DETAILS	
Assay	RP SARS-CoV-2

Assay	RP SARS-CoV-2	Cartridge SN	330412019	SN Operational module	20520098
	v3.1	Cartridge LOT	172044716	SN Analytical module	10520140
Sample	UTM	Expiration Date	2023-08-18	SW Version	1.4.0 build 5

Error None

Linear



Logarithmic

